Quality Assurance Certificate r-biophorm





RIDA®CUBE Ammonia **Product:**

REF Art. No.:	RCS4390
LOT Lot:	2004-1
	2021-06

Content	Lot	Expiry		
Reagent 1	2004-1	2021-06		
Reagent 2	2004-1	2021-06		
RFID Card	2004-1	2021-06		
IFU (and other accompanying documents where applicable)				

Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	100	92	92
Sample 2	20	19	95
Reagent blank (water sample)	0	2	/

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	25	24	96
Sample 2	5	5	100
Reagent blank (water sample)	0	0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2020-05-07



QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.