Quality Assurance Certificate r-biophorm®





Enzytec™ D-3-Hydroxybutyric acid **Product:**

REF	Art. No.:	E2610
LOT	Lot:	RBP003
\square	Expiry:	2022-09

Test	Specifications	Results
HBA low control	1.6 (1.2 – 2.0) mg/dL	1.85 – 1.82 mg/dL
HBA elevated control	40.1 (28.1 – 52.1) mg/dL	40.21 – 40.24 mg/dL
Reagent 1 visual check smell pH (25°C ± 1°C)	clear, colourless no characteristic odour 8.3 ± 0.1	conforms conforms 8.300
Reagent 2 visual check smell pH after reconstitution (25°C ± 1°C)	lyophilized powder no characteristic odour 8.0 ± 0.1	conforms conforms 8.000
Reagent 3 visual check smell pH (25°C ± 1°C)	colourless no characteristic odour N/A	conforms conforms
Reagent 4 visual check smell pH (25°C ± 1°C)	clear, colourless no characteristic odour N/A	conforms conforms

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2021-03-15



QAC-Version: 2021-03-16